

# Calibration Certificate

Calibration Laboratory accredited by the Swiss Accreditation Service  
(SAS) Accreditation number SCS 0032

**METTLER TOLEDO**



## Mettler-Toledo GmbH

8606 Greifensee, Switzerland

Phone: +41 44 944 23 15

Fax: +41 44 944 34 10

Email: calibration@mt.com

## Certificate

Certificate No.	T17682	Date of Calibration	2021-04-13
-----------------	--------	---------------------	------------

## Customer

Company	Mettler-Toledo Pac Rim AG
Address	TW-114 Taipei

## Summary

Manufacturer	METTLER TOLEDO
Object	Pt1000 temperature sensor
Calibration	Temperature
Order number	144413163
Identification number	TH191
Comments	Recalibration

Technical manager

  
Michael Kässer

Greifensee, 2021-04-16

Metrologist

  
Damir Tomas

**Object**

Manufacturer	METTLER TOLEDO
Object	Pt1000 temperature sensor
Identification number	TH191

**Calibration**

Type	Pt1000 temperature sensor
Calibration	Temperature
Condition at Calibration	Used
Comments	Recalibration

**Ambient Conditions**

Not relevant.

**Result**

Ref. Temp. ° C	Indication (Object) ° C	Deviation K	Uncertainty K	Tolerance K	In tol.?
50.03	50.09	+0.06	0.15	± 0.55	yes
100.97	100.80	-0.17	0.15	± 0.8	yes
160.75	160.25	-0.50	0.15	± 1.1	yes

**Calibration Method**

The temperature sensor and reference sensor are compared in a climate chamber in an equalizing aluminum block. The resistance of the temperature sensor is measured at different temperatures and converted into a temperature value using standard coefficients (EN 60751).

This calibration certificate documents the traceability to national standards, which realize the physical units of measurement (SI).

**Conformity Statement**

The deviation fulfills (or doesn't fulfill) the tolerance limit of EN 60751:2008, class B, as indicated in the last column of the table.

**SAS**

Calibration Laboratory accredited by the Swiss Accreditation Service (SAS). Accreditation number SCS 0032.

The Swiss Accreditation Service (SAS) is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates.

**Remarks**

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k = 2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%. This calibration certificate documents the traceability to national standards, which realize the physical units of measurement (SI).