

Address 900 Middlesex Turnpike, Bldg. 8, Billerica, MA 01821, USA
 Telephone +1 781-301-8600
 Fax +1 781-301-8701
 Internet www.mt.com/thornton

P/N: 58 082 305
Certificate of Test and Accuracy

Mettler-Toledo Thornton, Inc. hereby certifies that the following instrument performs within specification. The instrument has passed our quality control Certification of Calibration standards and has been calibrated using test standards traceable to the National Institute of Standards and Technology (NIST). Annual calibration is recommended.

Customer:	Mettler-Toledo Pac Rim AG	Date Performed:	08-Feb-2024
Manufacturer:	Mettler-Toledo Thornton, Inc	Technician:	BD
Part Number:	58 082 305	Relative Humidity:	23 %
Serial Number:	775285		
Room Temperature:	19 °C		
Calibration Procedure:	TP58170011		

Calibration Standards

Digital Multi-Meter:	Make:	Model:	S/N:	Stds. Cal.Due:
	HP	34401A	US36079097	31-Dec-2024

As Found Data:

Conductivity Test Results:

Nominal (ohms):	Label*(ohms):	Measured*(ohms):	Deviation (%):	Limit (+/-%):	Status:
4.0 M	3.9750 M	3.9744 M	-0.02	0.08	PASS
1.0 M	0.99964 M	0.9997 M	0.00	0.08	PASS
100.0 K	100.04 K	100.0496 K	0.01	0.08	PASS
10.0 K	10.003 K	10.0028 K	0.00	0.08	PASS
1.0 K	1.0001 K	1.0002 K	0.00	0.08	PASS
10.1	10.175	10.1677	0.01	0.03	PASS

After Calibration Test Data:

Conductivity Test Results:

Nominal (ohms):	Label*(ohms):	Measured*(ohms):	Deviation (%):	Limit (+/-%):	Status:
4.0 M	3.9744 M	3.9744 M	0.00	0.08	PASS
1.0 M	0.99965 M	0.9997 M	0.00	0.08	PASS
100.0 K	100.05 K	100.0496 K	0.00	0.08	PASS
10.0 K	10.003 K	10.0028 K	0.00	0.08	PASS
1.0 K	1.0002 K	1.0002 K	0.00	0.08	PASS
10.1	10.168	10.1677	0.00 ohms	0.03 ohms	PASS

* Values on the calibrator label may be rounded to fewer digits than the measured values.