

# Calibration Certificate

Calibration Laboratory accredited by the Swiss Accreditation Service  
(SAS) Accreditation number SCS 0032

**METTLER TOLEDO**



**Mettler-Toledo GmbH**

8606 Greifensee, Switzerland

Phone: +41 44 944 23 15

Fax: +41 44 944 34 10

Email: calibration@mt.com

## Certificate

Certificate No.	T19330	Date of Calibration	2022-07-04
-----------------	--------	---------------------	------------

## Customer

Company	Mettler-Toledo Pac Rim AG
Address	TW-114 Taipei

## Summary

Manufacturer	METTLER TOLEDO
Object	Pt1000 temperature sensor
Calibration	Temperature
Order number	145114322
Identification number	TH034, SVC-B023
Comments	Recalibration

Head of Calibration Lab

Luca Büchi

Greifensee, 2022-07-06

Metrologist

Esther Blöchlinger

**Object**

Manufacturer	METTLER TOLEDO
Object	Pt1000 temperature sensor
Identification number	TH034, SVC-B023

**Calibration**

Type	Pt1000 temperature sensor
Calibration	Temperature
Condition at Calibration	Used
Comments	Recalibration

**Ambient Conditions**

Not relevant.

**Result**

Ref. Temp. °C	Indication (Object) °C	Deviation K	Uncertainty K	Tolerance K	In tol.?
49.82	49.83	+0.01	0.15	± 0.55	yes
100.5	100.46	-0.04	0.15	± 0.8	yes
159.84	159.67	-0.17	0.15	± 1.1	yes

**Calibration Method**

The temperature sensor and reference sensor are compared in a climate chamber in an equalizing aluminum block. The resistance of the temperature sensor is measured at different temperatures and converted into a temperature value using standard coefficients (EN 60751).

This calibration certificate documents the traceability to national standards, which realize the physical units of measurement (SI).

**Conformity Statement**

The deviation fulfills (or doesn't fulfill) the tolerance limit of EN 60751:2008, class B, as indicated in the last column of the table.

**SAS**

Calibration Laboratory accredited by the Swiss Accreditation Service (SAS). Accreditation number SCS 0032.

The Swiss Accreditation Service (SAS) is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates.

**Remarks**

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k = 2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%. This calibration certificate documents the traceability to national standards, which realize the physical units of measurement (SI).