

B031

Calibration Certificate

Calibration Laboratory accredited by the Swiss Accreditation Service
(SAS) Accreditation number SCS 0032

METTLER TOLEDO



Mettler-Toledo GmbH
8606 Greifensee, Switzerland
Phone: +41 44 944 23 15
Fax: +41 44 944 34 10
Email: calibration@mt.com

Certificate

Certificate No.	T20360	Date of Calibration	2023-04-03
-----------------	--------	---------------------	------------

Customer

Company	Mettler-Toledo Pac Rim AG
Address	TW-114 Taipei

Summary

Manufacturer	METTLER TOLEDO
Object	Pt1000 temperature sensor
Calibration	Temperature
Order number	145557837
Identification number	TJ686
Comments	Initial calibration

財編	SVC-B031
品名	HX/HS Temperature Adjustment kit)
S/N	TJ686

Head of Calibration Lab

Marc Scheurmann

Metrologist

Esther Blöchlinger

Greifensee, 2023-04-20

Object

Manufacturer	METTLER TOLEDO
Object	Pt1000 temperature sensor
Identification number	TJ686

Calibration

Type	Pt1000 temperature sensor
Calibration	Temperature
Condition at Calibration	New
Comments	Initial calibration

Ambient Conditions

Not relevant.

Result

Ref. Temp. ° C	Indication (Object) ° C	Deviation K	Uncertainty K	Tolerance K	In tol.?
49.85	49.94	+0.09	0.15	± 0.55	yes
100.53	100.54	+0.01	0.15	± 0.8	yes
160.03	159.86	-0.17	0.15	± 1.1	yes

Calibration Method

The temperature sensor and reference sensor are compared in a climate chamber in an equalizing aluminum block. The resistance of the temperature sensor is measured at different temperatures and converted into a temperature value using standard coefficients (EN 60751).

This calibration certificate documents the traceability to national standards, which realize the physical units of measurement (SI).

Conformity Statement

The deviation fulfills (or doesn't fulfill) the tolerance limit of EN 60751:2008, class B, as indicated in the last column of the table.

SAS

Calibration Laboratory accredited by the Swiss Accreditation Service (SAS). Accreditation number SCS 0032.

The Swiss Accreditation Service (SAS) is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates.

Remarks

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 2$, which for a normal distribution corresponds to a coverage probability of approximately 95%. This calibration certificate documents the traceability to national standards, which realize the physical units of measurement (SI).