



Rpt No. : EKR2315126

第2頁 共2頁

校正結果：(absolute)

項次	器示值		標準值	器差值	擴充不確定度
	kPa	hPa	hPa	hPa	hPa
1	93.04	930.4	930.0	0.4	0.4
2	96.04	960.4	960.0	0.4	0.4
3	99.04	990.4	990.0	0.4	0.4
4	102.04	1020.4	1020.0	0.4	0.4
5	105.04	1050.4	1050.0	0.4	0.4

校正說明：

1. 器差值 = 器示值 - 標準值
2. 器示值係指校正件經重覆量測(升、降壓)三次共計六次之量測平均值
3. 標準值指工作標準件係經追溯後修正值
4. 擴充不確定度約以 95 %信賴水準, $k = 2$ 表示
5. 校正程序: CALP-EK-P003(V 2.7) 氣壓壓力(錶、計) 校正作業程序
6. 單位轉換: 1 hPa = 0.1 kPa

---THE END---

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com.tw/Terms-and-Conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com.tw/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.